Brazoria County Crime Lab

Toxicology Section

Headspace GC-FID Maintenance Log

Instrument: BAC-2 **Month/Year:** March 2025

Date	Nitrogen Tank Pressure	Hydrogen Tank Pressure	Helium Tank Pressure	Air Control Pass/Fail	PC Pass/Fail	Comments	Analyst Initials
3/10/2025	1900	700	2500	Pass	Pass	N2 tank changed 3/6/2025. He tank changed 3/10/2025.	mh

Signature: Kayla M. Baylor Digitally signed by Kayla M. Baylor Date: 2025.06.19 10:16:26 -05'00' **Date:** 6/19/2025