Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: MPC2000

Monthly Check

Date: 3/4/2025 Temp: 21.5 °C
Analyst: DN Humidity: 66 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.01 g	1	5.00 g	5.00 g	1	30.00 g	30.01 g
2	1.00 g	1.00 g	2	5.00 g	5.01 g	2	30.00 g	30.01 g
3	1.00 g	1.00 g	3	5.00 g	5.00 g	3	30.00 g	30.01 g

Surrogate Checks

Date: 3/10/2025 Temperature: 21.0 °C Weight Set Used: 155617
Analyst: DN Humidity: 40 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.98 g
2 1.00 g	2 5.01 g	2 99.98 g
3 1.00 g	3 5.01 g	3 99.98 g

Date: 3/17/2025 Temperature: 21.1 °C Weight Set Used: 155617
Analyst: DN Humidity: 45 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.98 g
2 1.00 g	2 5.01 g	2 99.98 g
3 1.00 g	3 5.00 g	3 99.98 g

Date: 3/24/2025 Temperature: 21.1 °C Weight Set Used: 155617
Analyst: DN Humidity: 64 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.98 g
2 1.00 g	2 5.00 g	2 99.97 g
3 1.00 g	3 5.00 g	3 99.98 g

Date: 3/31/2025 Temperature: 21.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 70 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.97 g
2 1.00 g	2 5.00 g	2 99.97 g
3 1.00 g	3 5.01 g	3 99.98 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: AP210S

Monthly Check

Date: 3/4/2025 Temp: 21.5 °C
Analyst: DN Humidity: 66 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 0.9999 g	1 3.00 g	1 3.0000 g	1 5.00 g	1 5.0000 g
2 1.00 g	2 0.9999 g	2 3.00 g	2 3.0000 g	2 5.00 g	2 4.9999 g
3 1.00 g	3 1.0000 g	3 3.00 g	3 2.9999 g	3 5.00 g	3 4.9999 g

Surrogate Checks

Date: 3/10/2025 Temperature: 21.0 °C Weight Set Used: 155617
Analyst: DN Humidity: 40 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0051 g	1 99.9593 g
2 1.0009 g	2 5.0052 g	2 99.9593 g
3 1.0009 g	3 5.0051 g	3 99.9593 g

Date: 3/17/2025 Temperature: 21.1 °C Weight Set Used: 155617
Analyst: DN Humidity: 45 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0009 g	1 5.0052 g	1 99.9593 g
2 1.0010 g	2 5.0053 g	2 99.9593 g
3 1.0009 g	3 5.0052 g	3 99.9594 g

Date: 3/24/2025 Temperature: 21.1 °C Weight Set Used: 155617
Analyst: DN Humidity: 64 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0009 g	1 5.0051 g	1 99.9595 g
2 1.0009 g	2 5.0050 g	2 99.9593 g
3 1.0009 g	3 5.0052 g	3 99.9594 g

Date: 3/31/2025 Temperature: 21.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 70 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0009 g	1 5.0052 g	1 99.9595 g
2 1.0009 g	2 5.0051 g	2 99.9595 g
3 1.0009 g	3 5.0051 g	3 99.9595 g