Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: MPC2000

Monthly Check

Date: 8/1/2025 Temp: 20.6 °C
Analyst: KB Humidity: 70 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 0.99 g	1 5.00 g	1 4.99 g	1 30.00 g	1 30.00 g
2 1.00 g	2 1.00 g	2 5.00 g	2 5.01 g	2 30.00 g	2 30.01 g
3 1.00 g	3 1.01 g	3 5.00 g	3 5.01 g	3 30.00 g	3 30.00 g

Surrogate Checks

Date: 8/4/2025 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 73 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.96 g
2 1.00 g	2 5.01 g	2 99.95 g
3 1.01 g	3 5.00 g	3 99.96 g

Date: 8/11/2025 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 71 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.97 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.00 g	3 5.01 g	3 99.97 g

Date: 8/18/2025 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: DN Humidity: 75 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.97 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.01 g	3 5.01 g	3 99.97 g

Date: 8/25/2025 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: DN Humidity: 73 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.96 g
2 1.01 g	2 5.00 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: AP210S

Monthly Check

Date: 8/1/2025 Temp: 20.6 °C
Analyst: KB Humidity: 70 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 0.9999 g	1 3.00 g	1 2.9999 g	1 5.00 g	1 4.9998 g
2 1.00 g	2 0.9999 g	2 3.00 g	2 2.9999 g	2 5.00 g	2 4.9999 g
3 1.00 g	3 0.9999 g	3 3.00 g	3 2.9999 g	3 5.00 g	3 4.9999 g

Surrogate Checks

Date: 8/4/2025 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 73 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0009 g	1 5.0051 g	1 99.9591 g
2 1.0009 g	2 5.0051 g	2 99.9591 g
3 1.0009 g	3 5.0051 g	3 99.9590 g

Date: 8/11/2025 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 71 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0052 g	1 99.9592 g
2 1.0009 g	2 5.0052 g	2 99.9592 g
3 1.0009 g	3 5.0052 g	3 99.9592 g

Date: 8/18/2025 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: DN Humidity: 75 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0008 g	1 5.0052 g	1 99.9590 g
2 1.0010 g	2 5.0052 g	2 99.9590 g
3 1.0009 g	3 5.0052 g	3 99.9590 g

Date: 8/25/2025 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: DN Humidity: 73 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0009 g	1 5.0052 g	1 99.9589 g
2 1.0009 g	2 5.0051 g	2 99.9590 g
3 1.0009 g	3 5.0052 g	3 99.9589 g