Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: MPC2000

Monthly Check

Date: 11/3/2025 Temp: 20.5 °C
Analyst: KK/DN Humidity: 54 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.00 g	1	5.00 g	5.00 g	1	30.00 g	29.99 g
2	1.00 g	1.01 g	2	5.00 g	5.00 g	2	30.00 g	30.00 g
3	1.00 g	1.00 g	3	5.00 g	5.00 g	3	30.00 g	30.00 g

Surrogate Checks

Date: 11/10/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 32 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.01 g	Balance Readout 1: 99.98 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.01 g	Balance Readout 2: 99.98 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.00 g	Balance Readout 3: 99.96 g

Date: 11/17/2025 Temperature: 20.5 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 68 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.00 g	Balance Readout 1: 99.96 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.00 g	Balance Readout 2: 99.97 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.97 g

Date: 11/24/2025 Temperature: 20.1 °C Weight Set Used: 155617
Analyst: KK/KB Humidity: 74 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.01 g	Balance Readout 1: 99.97 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.01 g	Balance Readout 2: 99.97 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.00 g	Balance Readout 3: 99.96 g

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

N/A

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: g	Balance Readout 1: g	Balance Readout 1: g
Balance Readout 2: g	Balance Readout 2: g	Balance Readout 2: g
Balance Readout 3: g	Balance Readout 3: g	Balance Readout 3: g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: AP210S

Monthly Check

Date: 11/3/2025 Temp: 20.7 °C
Analyst: KK/DN Humidity: 54 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 0.9997 g	1 3.00 g	1 2.9999 g	1 5.00 g	1 4.9999 g
2 1.00 g	2 0.9999 g	2 3.00 g	2 2.9998 g	2 5.00 g	2 4.9999 g
3 1.00 g	3 0.9999 g	3 3.00 g	3 3.0000 g	3 5.00 g	3 4.9999 g

Surrogate Checks

Date: 11/10/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 32 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0007 g	1 5.0052 g	1 99.9587 g
2 1.0009 g	2 5.0052 g	2 99.9587 g
3 1.0009 g	3 5.0051 g	3 99.9586 g

Date: 11/17/2025 Temperature: 20.5 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 68 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0009 g	1 5.0052 g	1 99.9592 g
2 1.0009 g	2 5.0051 g	2 99.9590 g
3 1.0010 g	3 5.0051 g	3 99.9590 g

Date: 11/24/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: KK/KB Humidity: 73 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0009 g	1 5.0052 g	1 99.9589 g
2 1.0009 g	2 5.0052 g	2 99.9589 g
3 1.0009 g	3 5.0051 g	3 99.9588 g

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

N/A

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 g	1 g	1 g
2 g	2 g	2 g
3 g	3 g	3 g