Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: MPC2000

Monthly Check

Date: 2/2/2026 Temp: 22.1 °C
Analyst: KK Humidity: 30 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.00 g	1	5.00 g	5.00 g	1	30.00 g	30.00 g
2	1.00 g	1.00 g	2	5.00 g	5.00 g	2	30.00 g	30.01 g
3	1.00 g	1.00 g	3	5.00 g	5.00 g	3	30.00 g	30.01 g

Surrogate Checks

Date: 2/9/2026 Temperature: 21.0 °C Weight Set Used: 155617
Analyst: KK Humidity: 59 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.00 g	Balance Readout	1 5.01 g	Balance Readout	1 99.97 g
	2 1.00 g		2 5.00 g		2 99.98 g
	3 1.00 g		3 5.01 g		3 99.98 g

Date: 2/17/2026 Temperature: 21.7 °C Weight Set Used: 155617
Analyst: DN Humidity: 61 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.00 g	Balance Readout	1 5.01 g	Balance Readout	1 99.96 g
	2 1.00 g		2 5.00 g		2 99.96 g
	3 1.00 g		3 5.00 g		3 99.96 g

Date: 2/23/2026 Temperature: 21.0 °C Weight Set Used: 155617
Analyst: KK Humidity: 29 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.00 g	Balance Readout	1 5.01 g	Balance Readout	1 99.97 g
	2 1.00 g		2 5.00 g		2 99.97 g
	3 1.01 g		3 5.00 g		3 99.97 g

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 g	Balance Readout	1 g	Balance Readout	1 g
	2 g		2 g		2 g
	3 g		3 g		3 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification AP210S

Monthly Check

Date _____ Temp _____ °C
Analyst _____ Humidity _____ %
Weight Set Used _____

Notes: The balance remains out of service; therefore, no checks were performed this month. -DN 3/3/2026

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1 g	1	3.00 g	1 g	1	5.00 g	1 g
2	1.00 g	2 g	2	3.00 g	2 g	2	5.00 g	2 g
3	1.00 g	3 g	3	3.00 g	3 g	3	5.00 g	3 g

Surrogate Checks

-DN 3/3/2026

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 g	Balance Readout 1 g	Balance Readout 1 g
2 g	2 g	2 g
3 g	3 g	3 g

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 g	Balance Readout 1 g	Balance Readout 1 g
2 g	2 g	2 g
3 g	3 g	3 g

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 g	Balance Readout 1 g	Balance Readout 1 g
2 g	2 g	2 g
3 g	3 g	3 g

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 g	Balance Readout 1 g	Balance Readout 1 g
2 g	2 g	2 g
3 g	3 g	3 g