Pipette Monitoring and Maintenance

Manufacturer Transferpipette Model _____
Serial Number 22H33903 Capacity 2-20µl

Pipette Identification #3903

Date 4/15/2026 Analyst mh H_2O °C Start 20.3 H_2O °C End 20.3 Room °C 20.3 Humidity % 62 Theoretical ρ N/A

Low 10µl		Mid-Level 20µl		High		☑ Pass ☐ Re-Check ☐ Fail
Measured	Actual	Measured	Actual	Measured	Actual	Notes:
1. 10	1. 0.0102	1. 20	1. 0.0195	1.	1.	
2. 10	2. 0.0101	2. 20	2. 0.0199	2. 4/15/2026 mh	2.	
3. 10	3. 0.0101	3. 20	3. 0.0198	3.	3.	
Actual: $\bar{x}$ = 10.133 σ = 0.058 % 101.33		Actual: $\bar{x}$ = 19.733 σ = 0.208 % 98.67		Actual: $\bar{x}$ = _____ σ = _____ % _____		

Date _____ Analyst _____ H_2O °C Start _____ H_2O °C End _____ Room °C _____ Humidity % _____ Theoretical ρ _____

Low		Mid-Level		High		☐ Pass ☐ Re-Check ☐ Fail
Measured	Actual	Measured	Actual	Measured	Actual	Notes:
1.	1.	1.	1.	1.	1.	
2.	2.	2. 4/15/2026 mh	2.	2.	2.	
3.	3.	3.	3.	3.	3.	
Actual: $\bar{x}$ = _____ σ = _____ % _____		Actual: $\bar{x}$ = _____ σ = _____ % _____		Actual: $\bar{x}$ = _____ σ = _____ % _____		