Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: AX224N

Monthly Check

Date: 11/3/2025 Temp: 19.6 °C
Analyst: DN Humidity: 56 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.0000 g	1 3.00 g	1 3.0000 g	1 5.00 g	1 5.0000 g
2 1.00 g	2 1.0000 g	2 3.00 g	2 3.0000 g	2 5.00 g	2 5.0000 g
3 1.00 g	3 1.0000 g	3 3.00 g	3 3.0000 g	3 5.00 g	3 5.0001 g

Surrogate Checks

Date: 11/10/2025 Temperature: 19.3 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 33 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0053 g	1 99.9593 g
2 1.0008 g	2 5.0053 g	2 99.9593 g
3 1.0008 g	3 5.0053 g	3 99.9593 g

Date: 11/17/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 74 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0008 g	1 5.0053 g	1 99.9601 g
2 1.0008 g	2 5.0052 g	2 99.9600 g
3 1.0011 g	3 5.0049 g	3 99.6000 g

Date: 11/24/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: KK/KB Humidity: 71 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0052 g	1 99.9600 g
2 1.0009 g	2 5.0052 g	2 99.9600 g
3 1.0009 g	3 5.0052 g	3 99.9598 g

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

N/A

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 g	1 g	1 g
2 g	2 g	2 g
3 g	3 g	3 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD1

Monthly Check

Date: 11/3/2025 Temp: 19.6 °C
Analyst: KK/DN Humidity: 56 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.00 g	1	5.00 g	5.01 g	1	30.00 g	30.00 g
2	1.00 g	1.01 g	2	5.00 g	4.99 g	2	30.00 g	30.00 g
3	1.00 g	1.00 g	3	5.00 g	5.00 g	3	30.00 g	30.00 g

Surrogate Checks

Date: 11/10/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 33 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 0.99 g	Balance Readout 1: 5.02 g	Balance Readout 1: 99.97 g
Balance Readout 2: 1.01 g	Balance Readout 2: 4.99 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.95 g

Date: 11/17/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 73 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.01 g	Balance Readout 1: 99.95 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.00 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.01 g	Balance Readout 3: 5.00 g	Balance Readout 3: 99.97 g

Date: 11/24/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: KK/KB Humidity: 71 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.00 g	Balance Readout 1: 99.96 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.00 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.00 g	Balance Readout 3: 99.96 g

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: g	Balance Readout 1: g	Balance Readout 1: g
Balance Readout 2: g	Balance Readout 2: g	Balance Readout 2: g
Balance Readout 3: g	Balance Readout 3: g	Balance Readout 3: g

N/A

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD2

Monthly Check

Date: 11/3/2025 Temp: 19.6 °C
Analyst: KK/DN Humidity: 56 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	0.98 g	1	5.00 g	5.00 g	1	30.00 g	30.00 g
2	1.00 g	1.00 g	2	5.00 g	5.00 g	2	30.00 g	30.02 g
3	1.00 g	1.00 g	3	5.00 g	5.01 g	3	30.00 g	29.98 g

Surrogate Checks

Date: 11/10/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 33 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.00 g	Balance Readout	1 5.00 g	Balance Readout	1 99.95 g
	2 1.02 g		2 4.99 g		2 99.97 g
	3 1.00 g		3 5.00 g		3 99.97 g

Date: 11/17/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 73 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.00 g	Balance Readout	1 5.00 g	Balance Readout	1 99.96 g
	2 1.00 g		2 5.01 g		2 99.96 g
	3 1.00 g		3 5.01 g		3 99.96 g

Date: 11/24/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: KK/KB Humidity: 71 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.00 g	Balance Readout	1 4.99 g	Balance Readout	1 99.96 g
	2 0.97 g		2 5.01 g		2 99.96 g
	3 1.01 g		3 5.01 g		3 99.95 g

N/A

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 g	Balance Readout	1 g	Balance Readout	1 g
	2 g		2 g		2 g
	3 g		3 g		3 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD3

Monthly Check

Date: 11/3/2025 Temp: 19.6 °C
Analyst: KK/DN Humidity: 56 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 0.97 g	1 5.00 g	1 5.00 g	1 30.00 g	1 30.01 g
2 1.00 g	2 0.99 g	2 5.00 g	2 5.00 g	2 30.00 g	2 30.01 g
3 1.00 g	3 1.00 g	3 5.00 g	3 5.00 g	3 30.00 g	3 30.01 g

Surrogate Checks

Date: 11/10/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 33 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.00 g	2 99.97 g
3 1.00 g	3 5.01 g	3 99.96 g

Date: 11/17/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 73 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 0.99 g	1 5.01 g	1 99.97 g
2 0.98 g	2 5.00 g	2 99.96 g
3 0.98 g	3 5.01 g	3 99.96 g

Date: 11/24/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: KK/KB Humidity: 71 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.01 g	1 5.01 g	1 99.96 g
2 1.00 g	2 5.00 g	2 99.96 g
3 1.00 g	3 5.01 g	3 99.98 g

N/A

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 g	1 g	1 g
2 g	2 g	2 g
3 g	3 g	3 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD4

Monthly Check

Date: 11/3/2025 Temp: 19.6 °C
Analyst: KK/DN Humidity: 56 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	0.99 g	1	5.00 g	5.00 g	1	30.00 g	30.01 g
2	1.00 g	1.03 g	2	5.00 g	5.00 g	2	30.00 g	30.01 g
3	1.00 g	0.99 g	3	5.00 g	4.98 g	3	30.00 g	29.98 g

Surrogate Checks

Date: 11/10/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 33 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.02 g	Balance Readout 1: 99.98 g
Balance Readout 2: 0.99 g	Balance Readout 2: 5.01 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.98 g

Date: 11/17/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 73 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.00 g	Balance Readout 1: 99.95 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.00 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.00 g	Balance Readout 3: 99.95 g

Date: 11/24/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: KK/KB Humidity: 71 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.00 g	Balance Readout 1: 99.96 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.00 g	Balance Readout 2: 99.95 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.96 g

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

N/A

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: g	Balance Readout 1: g	Balance Readout 1: g
Balance Readout 2: g	Balance Readout 2: g	Balance Readout 2: g
Balance Readout 3: g	Balance Readout 3: g	Balance Readout 3: g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: M-Toledo

Monthly Check

Date: 11/3/2025 Temp: 19.3 °C
Analyst: KK/DN Humidity: 59 %
Weight Set Used: B323405602

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	2.00 kg	2.00 kg	1	2.00 kg	2.00 kg	1	2.00 kg	2.00 kg
2	2.00 kg	2.00 kg	2	2.00 kg	2.00 kg	2	2.00 kg	2.00 kg
3	2.00 kg	2.00 kg	3	2.00 kg	2.00 kg	3	2.00 kg	2.00 kg

Surrogate Checks

Date: 11/10/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 33 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Date: 11/17/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 69 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Date: 11/24/2025 Temperature: 20.5 °C Weight Set Used: 155617
Analyst: KK/KB Humidity: 71 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

N/A

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 kg	Balance Readout	1 kg	Balance Readout	1 kg
	2 kg		2 kg		2 kg
	3 kg		3 kg		3 kg

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: PJ6000

Monthly Check

Date: 11/3/2025 Temp: 19.9 °C
Analyst: KK/DN Humidity: 57 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.0 g	1	5.00 g	5.0 g	1	30.00 g	30.0 g
2	1.00 g	1.0 g	2	5.00 g	5.0 g	2	30.00 g	30.0 g
3	1.00 g	1.0 g	3	5.00 g	5.0 g	3	30.00 g	30.0 g

Surrogate Checks

Date: 11/10/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 32 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.0 g	Balance Readout 1: 5.0 g	Balance Readout 1: 99.9 g
Balance Readout 2: 1.0 g	Balance Readout 2: 5.0 g	Balance Readout 2: 100.0 g
Balance Readout 3: 1.0 g	Balance Readout 3: 5.0 g	Balance Readout 3: 100.0 g

Date: 11/17/2025 Temperature: 19.5 °C Weight Set Used: 155617
Analyst: KK/DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.0 g	Balance Readout 1: 5.0 g	Balance Readout 1: 99.9 g
Balance Readout 2: 1.0 g	Balance Readout 2: 5.0 g	Balance Readout 2: 99.9 g
Balance Readout 3: 1.0 g	Balance Readout 3: 5.0 g	Balance Readout 3: 100.0 g

Date: 11/24/2025 Temperature: 19.8 °C Weight Set Used: 155617
Analyst: KK/KB Humidity: 76 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.0 g	Balance Readout 1: 5.0 g	Balance Readout 1: 100.0 g
Balance Readout 2: 1.0 g	Balance Readout 2: 5.0 g	Balance Readout 2: 99.9 g
Balance Readout 3: 1.0 g	Balance Readout 3: 5.0 g	Balance Readout 3: 100.0 g

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: g	Balance Readout 1: g	Balance Readout 1: g
Balance Readout 2: g	Balance Readout 2: g	Balance Readout 2: g
Balance Readout 3: g	Balance Readout 3: g	Balance Readout 3: g

N/A