unity lab services

Thermo Electron N.A.
1400 Northpointe Parkway
West Palm Beach, FL 33407

Service report

Instrument:	NICOLET IS20 MID-IR DTGS	Ticket:	1244175
Serial number:	BEP2210424	Type of service:	Service Request
		Coverage type:	MW: Factory Equipment Warrnty
		Service technician:	Jared Stevens

Thank you for selecting us as your service provider.

Billing

BRAZORIA COUNTY CRIME LAB ACCOUNTS PAYABLE
3602 CR 45
ANGLETON 77515
Texas

Service location

BRAZORIA COUNTY CRIME LAB
3602 COUNTY RD CR45
ANGLETON 77515
Texas
Contact: DEREK SANDERS

Arrival date: Jan-18-2023
Completion date: Jan-05-2023

Warranty validity date: Aug-01-2022 - Aug-01-2023

Subject: VALPRO FAILING NOISE TEST
Issue: VALPRO FAILING NOISE TEST IN ALL REGIONS

FOUND LOW INTERFEROGRAM, REPLACE DETECTOR ELEMENT 714-229700

Quantity	Item #	Item# Description
1.0	714-229700	IS20 IS10 IZ10 DTGS KBR DET ELEMENT KIT

Actions taken:
Replaced the detector element and adjusted interferometer endplates to resolve all issues.

Unit operating as intended: Yes

© 2019 Thermo Fisher Scientific Inc. All rights reserved.
Any personal data will be processed in accordance with Thermo Fisher Scientific's privacy notice, available here:
https://www.thermofisher.com/privacy-policy.html

ThermoFisher SCIENTIFIC

unity lab services

Thermo Electron N.A.
1400 Northpointe Parkway
West Palm Beach, FL 33407


Customer

Derek Sanders
Lab Manager

Signature
Date: Jan-18-2023

Technician

Jared Stevens



Signature
Date: Jan-18-2023

© 2019 Thermo Fisher Scientific Inc. All rights reserved.
Any personal data will be processed in accordance with Thermo Fisher Scientific's privacy notice, available here:
https://www.thermofisher.com/privacy-policy.html