unity lab services

Thermo Electron N.A.
1400 Northpointe Parkway
West Palm Beach, FL 33407

Service report

Instrument: NICOLET IS20 MID-IR DTGS
Serial number: BEP2210424

Ticket: 1857375
Purchase order: DR0476207
Type of service: PM (Contract)
Coverage type: MC: Contract
Contract number: 40416920
Service technician: Jared Stevens

Thank you for selecting us as your service provider.

Service location

BRAZORIA COUNTY CRIME LAB
3602 COUNTY RD CR45
ANGLETON 77515
Texas
Contact: KAYLA BAYLOR

Arrival date: May-13-2024
Completion date: May-13-2024

Subject: PM EXT WARRANTY SERVICE PLAN NICOLET FT-IR
Issue: PM-0040416920-000010-000001

Quantity	Item #	Item# Description

Actions taken:
I performed a PM and instrument functioned correctly after maintenance was completed. PM included successfully running ValPro, replacing desiccant pack/indicator, running advanced diagnostics, checking power supply voltages, collecting all pertinent data from BDiag, inspecting all hardware for damage/wear, collecting a background/sample, and ensuring the client is performing a bench alignment on a regular basis. The PM checklist has been attached to the ticket.

Unit operating as intended: Yes

© 2019 Thermo Fisher Scientific Inc. All rights reserved.
Any personal data will be processed in accordance with Thermo Fisher Scientific's privacy notice, available here:
https://www.thermofisher.com/privacy-policy.html

ThermoFisher SCIENTIFIC

unity lab services

Thermo Electron N.A.
1400 Northpointe Parkway
West Palm Beach, FL 33407

Customer

Kayla Baylor
Chemist

Kayla M. Baylor

Signature
Date: May-13-2024

Technician

Jared Stevens

Signature
Date: May-13-2024

© 2019 Thermo Fisher Scientific Inc. All rights reserved.
Any personal data will be processed in accordance with Thermo Fisher Scientific's privacy notice, available here:
https://www.thermofisher.com/privacy-policy.html

ThermoFisher SCIENTIFIC