Allometrics
when measurements matter
1425 B Atlantis Dr, Webster, Tx 77598
(800) 528-2246 Fax (281) 474-5482
www.allometrics.com

ACCREDITED CERT #2039.01

Pipette Calibration Certificate
001300897

Customer: BRAZORIA COUNTY AUDITOR
3602 COUNTY ROAD 45

ANGLETON, TX 77515-9575
Work Order: 00150588
Contact: ALEIA WINTERS (aleiaw@brazoria-county.com)

Asset ID: 21F95252
Serial No.: 21F95252
Equipment Type: PIPETTE

Result: PASS
Cal Date: 03/26/2026
Next Cal Date: 3/26/2027
Performed At: Allometric's Laboratory
Performed By: MGONZALES

TEST ENVIRONMENT DATA

Start:	Air Temperature: 20.7 °C	Water Temperature: 19.99 °C	Humidity: 50.0 %	Barometric Pressure: 1017 hPa
Stop:	Air Temperature: 20.7 °C	Water Temperature: 19.99 °C	Humidity: 50.0 %	Barometric Pressure: 1017 hPa

Manufacturer: BRANDTECH
Model Number: BTS-20
Capacity: 20 µl
Type: A - Variable Volume
Channel: Single
Accuracy Tol.: ± 0.20 µl [1]
Precision Tol.: ± 0.10 µl [1]
z-factor: 1.00285 µl/mg

[1] ISO 8655 tolerances.

TEST DATA

	As Found Data			As Returned Data		
Test Volume	2 µl	10 µl	20 µl	2 µl	10 µl	20 µl
Sample 1 (mg)	1.90	10.01	19.82	1.90	10.01	19.82
Sample 2 (mg)	2.00	10.13	19.96	2.00	10.13	19.96
Sample 3 (mg)	1.90	9.81	19.99	1.90	9.81	19.99
Sample 4 (mg)	1.95	9.95	19.86	1.95	9.95	19.86
Sample 5 (mg)	1.96	9.95	19.98	1.96	9.95	19.98
Sample 6 (mg)	1.95	9.98	19.96	1.95	9.98	19.96
Sample 7 (mg)	1.98	9.98	19.84	1.98	9.98	19.84
Sample 8 (mg)	1.97	9.99	19.97	1.97	9.99	19.97
Sample 9 (mg)	1.86	9.95	19.98	1.86	9.95	19.98
Sample 10 (mg)	2.01	10.06	19.84	2.01	10.06	19.84
Mean (mg)	1.95	9.98	19.92	1.95	9.98	19.92
Precision (µl)	0.048	0.083	0.070	0.048	0.083	0.070
V20 (µl)	1.96	10.01	19.98	1.96	10.01	19.98
Error (µl)	-0.04	0.01	-0.02	-0.04	0.01	-0.02
Result	Pass	Pass	Pass	Pass	Pass	Pass
Uncertainty k=2	± 0.097 µl	± 0.17 µl	± 0.14 µl	± 0.097 µl	± 0.17 µl	± 0.14 µl

*Result: The decision rule used is Simple Acceptance as defined in ILAC G8 and ISO 8655-6.

TEST EQUIPMENT

Description	ID	Serial Number	Cal. Due Date
LAB ENVIRONMENTAL MONITOR	1251	73167	5/31/2027
WEIGHT SET / 10MG TO 50G	1616	401299390	5/31/2026
DIGITAL THERMOMETER	3013	192608473	4/30/2026
BALANCE / PIPETTE BALANCE	6050	T0102697	3/31/2026

Balance Model	Sensitivity
AND AD4212B-101	101g x.1mg/31x.01/5.1x.001

PROCEDURE

Procedure Name	Description	Revision Level	Revision Date
5-4WI06	PIPETTE CALIBRATION IAW ISO 8655-6	I	7/18/2025

Inspections and measurements were performed in accordance with ISO 8655-6:2002(E), and ISO 17025 where applicable. Pass/Fail status determination is derived from allowable tolerances established in ISO 8655 unless other tolerances are specifically requested by the client. If this is an accredited report, then our accredited logo will appear in the top right corner and the uncertainty associated with these measurements were determined using a coverage factor of k=2 and is not included in the determination of Pass/Fail status. This calibration is traceable to the SI through NIST or other NMI utilizing the equipment shown on the calibration certificate.

ACCEPTANCE OF THIS CERTIFICATION OR CALIBRATION DOCUMENT OR THE CONTINUED USE OF THE EQUIPMENT LISTED ON THIS DOCUMENT INDICATES AGREEMENT WITH ALLOMETRICS INC. TERMS AND CONDITIONS AS FOUND AT https://www.allometrics.com/terms-and-conditions/ AND AGREEMENT THAT THOSE CONDITIONS GOVERN ANY TRANSACTION RELATED TO THE CERTIFICATION, CALIBRATION, REPAIR OR USE OF ANY EQUIPMENT LISTED ON OR ASSOCIATED WITH THIS DOCUMENT.

COMMENTS

--------- end report ---------

Approved and Released by: Terry Baldwin, Quality Manager
Date of Issue: 03/26/2026

This report applies only to the item calibrated. This report may not be reproduced, except in full, without the written consent from Allometrics, Inc. © 2026