BRAZORIA COUNTY SHERIFF'S OFFICE
CRIME LABORATORY
3602 County Road 45, Angleton, TX 77515
(979) 864-2250 • CrimeLab@brazoriacountytx.gov

SEIZED DRUG MEASUREMENT UNCERTAINTY

Balance Case Number: MU SD2
Report Number: 0001
Report Date: 7/17/2024

Balance: Ohaus AX2202 (SD2)
SN: C144891582
Effective Dates: 4/24/2024 to 10/24/2025

NOTE:

This measurement uncertainty value is already in effect as of April 24, 2024. It is only now being populated due to the laboratory's switch to paperless.

RESULTS:

	u (k=2)	u (k=3)
Reported Uncertainty:	0.03	0.04

File Attachment(s):

- UoM 2024_Ohaus AX2202 SD2

Laboratory Policy:

All weight measurements shall be reported at the 95% coverage probability, k=2, rounded to the second decimal place to correlate to the balance resolution.

Analyst: Aleia Winters 18 July 2024