Temperature Log

Refrigerator / Freezer Mini-freezer
(0°C to 8°C) (0°C and below)

Thermometer Manufacturer _____

S/N 230218774

Control Company
Installed: 06/07/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T - 01/02	-		
W - 01/03	-17°C		DS
Th - 01/04			
F - 01/05			
M - 01/08			
T - 01/09			
W - 01/10			
Th - 01/11			
F - 01/12	-18°C		DS
M - 01/15			
T - 01/16	-18°C		DS
W - 01/17			
Th - 01/18			
F - 01/19			
M - 01/22			
T - 01/23			
W - 01/24			
Th - 01/25			
F - 01/26	-18°C	No maintenance - installation of Dickson temp monitor/official transfer to Dickson will occur at a later date.	DS
M - 01/29	-18°C		DS
T - 01/30			
W - 01/31			

Temperature Log

Refrigerator (0°C to 8°C) / Freezer (0°C and below) Bio Received Fridge

Thermometer Manufacturer _____

S/N 230218770

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T- 01/02			
W- 01/03	1°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	2°C		DS
M- 01/15			
T- 01/16	3°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	2°C	No maintenance - installation of Dickson temp monitor/official transfer to Dickson will occur at a later date.	DS
M- 01/29	2°C		DS
T- 01/30			
W- 01/31			

Temperature Log

Refrigerator (0°C to 8°C) / Freezer (0°C and below) Biofridge #7

Thermometer Manufacturer _____

S/N 230213764

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T- 01/02			
W- 01/03	1°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	3°C		DS
M- 01/15			
T- 01/16	3°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	3°C	No maintenance - installation of Dickson temp monitor/official transfer to Dickson will occur at a later date.	DS
M- 01/29	4°C		DS
T- 01/30			
W- 01/31			

Temperature Log

Refrigerator / Freezer Bio Fridge # 6
(0°C to 8°C) (0°C and below)

Thermometer Manufacturer _

S/N 230218771

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T- 01/02			
W- 01/03	2°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	2°C		DS
M- 01/15			
T- 01/16	3°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	3°C		DS
M- 01/29	2°C		DS
T- 01/30			
W- 01/31			

Temperature Log

Refrigerator (0°C to 8°C) / Freezer (0°C and below) Bio Fridge #5

Thermometer Manufacturer _

S/N 230218767

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T- 01/02			
W- 01/03	2°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	4°C		DS
M- 01/15			
T- 01/16	3°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	3°C		DS
M- 01/29	3°C		DS
T- 01/30			
W- 01/31			

Temperature Log

Refrigerator (0°C to 8°C) / Freezer (0°C and below) Biofridge #2

Thermometer Manufacturer _____

S/N 230218765

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T- 01/02			
W- 01/03	2°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	3°C		DS
M- 01/15			
T- 01/16	5°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	3°C		DS
M- 01/29	2°C		DS
T- 01/30			
W- 01/31			

Temperature Log

Refrigerator / Freezer Biofridge #1
(0°C to 8°C) (0°C and below)

Thermometer Manufacturer _____

S/N 220218768

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T- 01/02			
W- 01/03	2°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	3°C		DS
M- 01/15			
T- 01/16	3°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	2°C		DS
M- 01/29	2°C		DS
T- 01/30			
W- 01/31			

Temperature Log

Refrigerator / ~~Freezer~~ Specimen Freezer #1
(0°C to 8°C) (0°C and below)

Thermometer Manufacturer _____

S/N 230218775

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Jan 2024

Date	Temperature	Maintenance	Analyst
T- 01/02			
W- 01/03	-18°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	-17°C		DS
M- 01/15			
T- 01/16	-16°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	16°C	Upon temp recording, the temp was elevated outside of its range. All evidence was removed from ~~refrig~~ freezer and transferred to ~~Sherman~~ freezer 2. There was a power outage on 01/24/2024, it is suspected that the outage may be the cause of the issue. DS	DS
M- 01/29	20°C	→ The unit was unplugged and plugged into different outlet during resolution of the issue. The unit is removed from service until repair or replacement is made DS	
T- 01/30			
W- 01/31			

Temperature Log

Refrigerator / ~~Freezer~~ Specimen Freezer * 2
(0°C to 8°C) (0°C and below)

Thermometer Manufacturer _

S/N 230218722

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T- 01/02			
W- 01/03	-14°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	-15°C		DS
M- 01/15			
T- 01/16	-15°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	-14°C		DS
M- 01/29	-15°C		DS
T- 01/30			
W- 01/31			

Temperature Log

Refrigerator / Freezer Specimen Refrigerator / Freezer #4
(0°C to 8°C) (0°C and below)

Thermometer Manufacturer _

S/N 230218761

Control Company
Installed: 07/24/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T. 01/02			
W. 01/03	3°C		DS
Th. 01/04			
F. 01/05			
M. 01/08			
T. 01/09			
W. 01/10			
Th. 01/11			
F. 01/12	3°C		DS
M. 01/15			
T. 01/16	4°C		DS
W. 01/17			
Th. 01/18			
F. 01/19			
M. 01/22			
T. 01/23			
W. 01/24			
Th. 01/25			
F. 01/26	4°C	No maintenance - installation of Dickson temp monitor/official transfer to Dickson will occur at a later date.	DS
M. 01/29	3°C		DS
T. 01/30			
W. 01/31			

Temperature Log

Refrigerator (0°C to 8°C) / Freezer (0°C and below) Bio Fridge #3

Thermometer Manufacturer _

S/N 230218762

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T- 01/02			
W- 01/03	1°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	1°C		DS
M- 01/15			
T- 01/16	1°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	1°C		DS
M- 01/29	1°C		DS
T- 01/30			
W- 01/31			

Temperature Log

Refrigerator / ~~Freezer~~ Standard Freezer
(0°C to 8°C) (0°C and below)

Thermometer Manufacturer _

S/N 230218749

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T- 01/02			
W- 01/03	-15°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	-13°C		DS
M- 01/15			
T- 01/16	-10°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	-15°C		DS
M- 01/29	-14°C		DS
T- 01/30			
W- 01/31			

Temperature Log

Refrigerator / Freezer Standards Fridge
(0°C to 8°C) (0°C and below)

Thermometer Manufacturer _____

S/N 230218766

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T- 01/02			
W- 01/03	3°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	4°C		DS
M- 01/15			
T- 01/16	2°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	3°C		DS
M- 01/29	3°C		DS
T- 01/30			
W- 01/31			

Temperature Log

Refrigerator (0°C to 8°C) / Freezer (0°C and below) Overflow Fridge

Thermometer Manufacturer _____

S/N 230218763

Control Company
Installed: 06/08/2023
Cal Date: 03/30/2025

Date	Temperature	Maintenance	Analyst
Jan 2024			
T- 01/02			
W- 01/03	1°C		DS
Th- 01/04			
F- 01/05			
M- 01/08			
T- 01/09			
W- 01/10			
Th- 01/11			
F- 01/12	1°C		DS
M- 01/15			
T- 01/16	1°C		DS
W- 01/17			
Th- 01/18			
F- 01/19			
M- 01/22			
T- 01/23			
W- 01/24			
Th- 01/25			
F- 01/26	1°C	No maintenance - installation of Dickson temp monitor / official transfer to Dickson will occur at a later date.	DS
M- 01/29	1°C		DS
T- 01/30			
W- 01/31			