



# Pipette Monitoring and Maintenance

Manufacturer Gilson Model micromach  
 Serial Number 5050 Capacity 100-1000ul

## Pipette Identification \_\_\_\_\_

Date <u>4/15/2024</u> Analyst <u>mh</u> H <sub>2</sub> O °C Start <u>21.0</u> H <sub>2</sub> O °C End <u>21.0</u> Room °C <u>21.0</u> Humidity % <u>62</u> Theoretical p <u>n/a</u>							
Low <u>100ul</u>		Mid-Level <u>500ul</u>		High <u>1000ul</u>		<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Re-Check <input type="checkbox"/> Fail <div style="border: 1px solid black; padding: 5px; min-height: 40px;">Notes:</div>	
Measured 1. <u>0.1000</u> 2. <u>0.1000</u> 3. <u>0.1000</u>	Actual 1. <u>0.0952</u> 2. <u>0.0959</u> 3. <u>0.0968</u>	Measured 1. <u>0.5000</u> 2. <u>0.5000</u> 3. <u>0.5000</u>	Actual 1. <u>0.4913</u> 2. <u>0.4942</u> 3. <u>0.5011</u>	Measured 1. <u>1.0000</u> 2. <u>1.0000</u> 3. <u>1.0000</u>	Actual 1. <u>0.9939</u> 2. <u>0.9921</u> 3. <u>0.9919</u>		

Date _____ Analyst _____ H <sub>2</sub> O °C Start _____ H <sub>2</sub> O °C End _____ Room °C _____ Humidity % _____ Theoretical p _____							
Low		Mid-Level		High		<input type="checkbox"/> Pass <input type="checkbox"/> Re-Check <input type="checkbox"/> Fail <div style="border: 1px solid black; padding: 5px; min-height: 40px;">Notes:</div>	
Measured 1. _____ 2. _____ 3. _____	Actual 1. _____ 2. _____ 3. _____	Measured 1. _____ 2. _____ 3. _____	Actual 1. _____ 2. _____ 3. _____	Measured 1. _____ 2. _____ 3. _____	Actual 1. _____ 2. _____ 3. _____		